

FEBRA - Standard Cuff Nitrile Gloves - Powder Free Specification Sheet

Spec No. FEB024A	PRODUCT SPECIFICATION		Revision: 2 Effective: 13/03/2020
Product Information:			
Trade Name	Non-sterile Violet Powder Free Nitrile Examination Gloves - FEBRA - Standard Cuff		
Intended Use	A patient examination glove is a disposable device intended for medical purpose that is worn on the examiner's hand or finger to prevent contamination between patient and examiner		
Product Conformance	EN455 Medical Use, EN374 Viral Penetration, EN420		
Material	Nitrile (Carboxylated Butadiene Acrylonitrile)		
Cuff Finishing	Beaded		
Color	Violet/Purple		
Type	Ambidextrous		
Surface finish	Full textured		
Product Specification:			
Dimension	Size	Palm Width (mm)	Length (mm)
	X-Small	80 ± 5mm	240min
	Small	85 ± 5mm	240min
	Medium	95 ± 5mm	240min
	Large	106 ± 5mm	240min
	Extra Large	115 ± 5mm	240min
Thickness	Single wall thickness (mm)		
		Specification	
	Cuff	0.10-0.13	
	Palm	0.12-0.15	
	Finger	0.16-0.22	
	Beading	0.5min	
Physical Properties		Un-aged	Aged
		Spec	Spec
	Force at Break (N)	6 min	6 min
Residual Powder	2.0 mg/glove maximum		
Quality Specification * AND – Accept No Defects	Inspection Criteria		Sampling Plan/AQL
	Dimension		13 pcs (Median)
	Force at Break		13 pcs (Median)
	Water Tight Test		G1 – AQL 1.5 (Single Normal Sampling)
	Visual Inspection – Major		G1 – AQL 2.5
	Visual Inspection – Minor		G1 – AQL 4.0
	Residual Powder		n = 5 (average)
	Packaging Labeling		n = 5 (AND)
	Glove Physical Counting		n = 5 (AND)
Packaging Configuration and Labeling			
Packing	100 Gloves per Inner Dispenser (by weight) 10 Dispensers per Outer Carton		
Labeling	Brand: Febra Made in Malaysia Single Use Marking/Non Sterile Marking		
Expiry Date	5 Years		



Dairy Spares Ltd

Godwin House, Mullbry Business Park, Shakespeare Way, Whitchurch, Shropshire, SY13 1LJ, UK
T: +44 (0)1948 667676 • E: post@dairyspares.co.uk • www.dairyspares.co.uk